



測定値 (mg/L)	1回目	0.543	0.812	0.979	0.980	0.967	0.974	0.988	0.979	1.013	1.003	1.009	0.997	1.008	1.011	1.016	1.016	1.017	1.039	1.040	0.976	1.032	1.039	1.043	1.020	1.059	1.060	1.074	1.092	1.109	1.094		
	2回目	0.495	0.796	0.939	0.969	0.983	0.996	0.985	0.998	1.003	1.002	1.016	0.999	1.020	1.010	1.019	1.018	0.992	1.018	1.082	1.032	1.018	1.039	1.063	1.053	1.056	1.070	1.085	1.118	1.199			
	3回目	0.475	0.788	0.983	0.954	0.976	0.980	0.985	1.006	0.998	1.004	1.009	1.011	1.027	1.008	1.015	1.017	1.017	1.045	1.025	1.027	1.032	1.056	1.040	1.082	1.055	1.058	1.090	1.086	1.119	1.121		
平均値(mg/L)①		0.504	0.799	0.967	0.968	0.971	0.979	0.990	0.990	1.003	1.003	1.007	1.008	1.011	1.013	1.014	1.017	1.017	1.025	1.028	1.028	1.032	1.038	1.041	1.055	1.056	1.058	1.078	1.088	1.115	1.138		
標準偏差②		0.035	0.012	0.024	0.013	0.005	0.005	0.006	0.014	0.009	0.001	0.004	0.010	0.014	0.006	0.003	0.002	0.001	0.029	0.011	0.053	0.000	0.019	0.002	0.032	0.003	0.002	0.011	0.004	0.006	0.055		
CV(室内)(%) (②/①)		6.930	1.530	2.516	1.349	0.487	0.468	0.575	1.432	0.863	0.058	0.379	0.977	1.413	0.616	0.317	0.150	0.057	2.831	1.094	5.155	0.000	1.834	0.200	3.011	0.289	0.189	0.982	0.348	0.494	4.791		
参加者平均との差		-0.522	-0.228	-0.059	-0.059	-0.056	-0.047	-0.037	-0.036	-0.023	-0.023	-0.020	-0.018	-0.015	-0.013	-0.013	-0.009	-0.009	-0.001	0.001	0.002	0.006	0.011	0.014	0.029	0.029	0.032	0.052	0.061	0.089	0.112		
Zスコア		-12.46	-5.44	-1.42	-1.40	-1.33	-1.13	-0.88	-0.87	-0.56	-0.55	-0.47	-0.44	-0.36	-0.32	-0.30	-0.21	-0.21	-0.02	0.03	0.05	0.14	0.27	0.34	0.68	0.70	0.76	1.23	1.46	2.13	2.67		
Grubss検定による棄却		○	○																														

参加者平均	1.026
標準偏差	0.042
CV(室内)%	1.02

Zスコア	事業所数	30
$ Z \leq 2$		26
$2 < Z < 3$		2
$3 \leq Z $		2