



測定値 (mg/L)	1回目	24.2	42.5	46.6	44.5	53.3	45.1	47.0	49.8	54.3	51.8	53.6	53.1	55.9	58.3	57.7	59.0	59.3	59.7	59.7	60.8	61.1	60.5	62.3	61.6	64.1	63.0	64.0	65.1	68.1	70.3	73.1	70.9	97.2	
	2回目	24.1	42.0	40.9	44.6	42.3	45.9	45.0	51.6	55.0	52.3	53.6	53.2	55.0	54.5	57.8	60.1	58.4	58.7	60.3	60.1	61.4	61.8	61.7	62.8	62.3	63.9	63.6	67.5	66.9	71.0	71.8	73.1	101.2	
	3回目	24.8	41.9	45.9	44.6	40.1	45.9	44.9	52.6	47.5	53.4	54.0	55.2	55.8	55.1	56.7	58.1	59.7	59.7	60.0	59.7	60.4	60.1	61.6	60.5	61.6	60.5	60.3	62.6	65.5	69.5	71.8	71.8	75.1	100.2
平均値(mg/L)①	24.4	42.1	44.5	44.6	45.2	45.6	45.6	51.3	52.3	52.5	53.7	53.8	55.6	56.0	57.4	59.1	59.1	59.5	59.9	60.4	60.9	61.3	61.5	62.0	62.3	62.4	63.4	66.0	68.2	71.0	72.2	73.1	99.5		
標準偏差②	0.38	0.32	3.11	0.06	7.07	0.46	1.18	1.42	4.14	0.82	0.23	1.18	0.49	2.04	0.61	4.14	1.00	0.67	0.68	0.35	0.35	0.68	0.70	0.92	0.69	1.80	1.87	0.72	1.29	1.30	0.75	0.75	2.12	2.08	
CV(室内)(%) (②/①)	1.55	0.76	6.99	0.13	15.64	1.01	2.60	2.76	7.93	1.56	0.43	2.20	0.89	3.65	1.06	1.70	1.13	1.14	0.58	0.58	3.12	1.14	1.49	1.12	2.89	3.00	1.14	1.95	1.91	1.06	1.04	2.90	2.09		
参加者平均との差	-33.14	-15.37	-13.04	-12.94	-12.27	-11.87	-11.87	-6.17	-5.24	-5.00	-3.77	-3.67	-1.94	-1.54	-0.10	1.56	1.63	1.96	2.40	2.93	3.36	3.80	4.00	4.50	4.80	4.90	5.90	8.53	10.66	13.53	14.73	15.56	42.03		
Zスコア	-3.94	-1.83	-1.55	-1.54	-1.46	-1.41	-1.41	-0.73	-0.62	-0.59	-0.45	-0.44	-0.23	-0.18	-0.01	0.19	0.19	0.23	0.29	0.35	0.40	0.45	0.48	0.53	0.57	0.58	0.70	1.01	1.27	1.61	1.75	1.85	4.99		
Grubbs検定による棄却	○																																		○

参加者平均	57.5
標準偏差	8.42
CV(室内)%	2.37

Zスコア	事業所数	33
$ Z \leq 2$		31
$2 < Z < 3$		0
$3 \leq Z $		2