



測定値 (mg/L)	1回目	3.100	4.500	4.600	5.200	5.300	5.400	5.600	5.500	5.400	5.710	5.500	5.600	5.600	5.700	5.700	5.800	5.800	5.800	5.900	5.900	5.900	5.700	6.100	5.700	6.000	6.100	6.200	6.100	6.300
2回目	3.300	4.600	4.500	5.100	5.400	5.400	5.300	5.300	5.400	5.500	5.600	5.600	5.600	5.600	5.800	5.700	5.800	5.800	5.900	5.900	5.900	5.900	5.900	5.800	5.900	5.800	6.000	6.100	6.300	6.300
3回目	3.300	4.500	4.500	5.200	5.100	5.300	5.300	5.500	5.500	5.500	5.400	5.500	5.600	5.600	5.600	5.700	5.800	5.900	5.800	5.900	5.900	5.900	6.100	5.900	6.200	6.000	6.000	6.000	6.100	6.200
平均値(mg/L)①	3.230	4.530	4.530	5.160	5.260	5.360	5.400	5.430	5.430	5.530	5.530	5.600	5.600	5.700	5.700	5.800	5.830	5.830	5.900	5.900	5.900	5.900	6.100	5.900	6.200	6.000	6.030	6.100	6.160	6.260
標準偏差②	0.116	0.058	0.058	0.058	0.153	0.058	0.173	0.116	0.058	0.158	0.058	0.000	0.000	0.100	0.000	0.000	0.058	0.058	0.000	0.000	0.000	0.000	0.200	0.153	0.252	0.116	0.058	0.100	0.116	0.058
CV(室内)(%)②/①	3.576	1.274	1.274	1.118	2.905	1.076	3.207	2.127	1.063	2.861	1.043	0.000	0.000	1.754	0.000	0.000	0.990	0.990	0.000	0.000	0.000	0.000	3.390	2.577	4.245	1.948	0.957	1.639	1.875	0.922
参加者平均との差	-2.419	-1.119	-1.119	-0.489	-0.389	-0.289	-0.249	-0.219	-0.219	-0.119	-0.119	-0.049	-0.049	0.051	0.051	0.151	0.181	0.181	0.251	0.251	0.251	0.251	0.251	0.281	0.281	0.281	0.381	0.451	0.511	0.611
Zスコア	-5.78	-2.67	-2.67	-1.17	-0.93	-0.69	-0.59	-0.52	-0.52	-0.28	-0.28	-0.12	-0.12	0.12	0.12	0.36	0.43	0.43	0.60	0.60	0.60	0.60	0.60	0.67	0.67	0.67	0.91	1.08	1.22	1.46
Grubss検定による棄却	○																													

参加者平均	5.65
標準偏差	0.42
CV(室内)%	1.40

Zスコア	事業所数	29
$ Z \leq 2$		26
$2 < Z < 3$		2
$3 \leq Z $		1