



測定値 (mg/L)	1回目	0.0051	0.0037	0.0039	0.0044	0.0037	0.0039	0.0039	0.0038	0.0036	0.0039	0.0040	0.0040	0.0037	0.0039	0.0038	0.0039	0.0040	0.0037	0.0039	0.0040	0.0036	0.0047	0.0039	0.0040	0.0037	0.0039	0.0036	0.0040	0.0036	0.0040	0.0041	0.0032	0.0037	0.0041
2回目	0.0051	0.0037	0.0039	0.0044	0.0037	0.0039	0.0039	0.0038	0.0038	0.0039	0.0040	0.0040	0.0038	0.0040	0.0037	0.0039	0.0039	0.0037	0.0038	0.0039	0.0036	0.0047	0.0040	0.0039	0.0042	0.0039	0.0036	0.0040	0.0036	0.0040	0.0036	0.0035	0.0037	0.0041	
3回目	0.0050	0.0037	0.0039	0.0043	0.0037	0.0039	0.0039	0.0038	0.0038	0.0039	0.0040	0.0040	0.0039	0.0037	0.0038	0.0038	0.0039	0.0037	0.0038	0.0039	0.0036	0.0046	0.0039	0.0039	0.0038	0.0039	0.0036	0.0040	0.0045	0.0034	0.0038	0.0041			
平均値(mg/L)(1)	0.00507	0.00370	0.00390	0.00437	0.00370	0.00390	0.00390	0.00380	0.00380	0.00393	0.00400	0.00397	0.00373	0.00400	0.00377	0.00387	0.00393	0.00370	0.00390	0.00390	0.00360	0.00467	0.00393	0.00393	0.00390	0.00390	0.00360	0.00400	0.00407	0.00337	0.00373	0.00410			
標準偏差(2)	0.000577	0.0000000	0.0000000	0.000577	0.0000000	0.0000000	0.0000000	0.0000000	0.0002000	0.000577	0.0000000	0.000577	0.0001000	0.000577	0.000577	0.000577	0.000577	0.0000000	0.0001000	0.0001000	0.0000000	0.000577	0.000577	0.000577	0.0002646	0.0000000	0.0000000	0.0000000	0.0004509	0.0001528	0.000577	0.0000000			
CV(室内)(%) (2/1)	1.14	0.00	0.00	1.32	0.00	0.00	0.00	0.00	5.26	1.47	0.00	1.45	1.55	2.50	1.53	1.49	1.47	0.00	2.56	2.56	0.00	1.24	1.47	1.47	6.79	0.00	0.00	0.00	11.08	4.53	1.55	0.00			
参加者平均との差	0.00121	-0.00016	0.00004	0.00051	-0.00016	0.00004	0.00004	-0.00006	-0.00006	0.00007	0.00014	0.00011	-0.00013	0.00014	-0.00009	0.00001	0.00007	-0.00016	0.00004	0.00004	-0.00026	0.00081	0.00007	0.00007	0.00004	0.00004	-0.00026	0.00014	0.00021	-0.00049	-0.00013	0.00024			
Zスコア	6.834	-0.877	0.219	2.796	-0.877	0.219	0.219	-0.329	-0.329	0.384	0.768	0.603	-0.713	0.768	-0.493	0.055	0.384	-0.877	0.219	0.219	-1.425	4.441	0.384	0.384	0.219	0.219	-1.425	0.768	1.151	-2.686	-0.713	1.316			
Grubbs検定による棄却	○																					○													

平均値	0.00386
標準偏差	0.000182
CV(室内)%	1.67

Zスコア	事業所数	32
Z  ≤ 2		28
2 <  Z  < 3		2
3 ≤  Z		2