



測定値 (mg/L)	1回目	0.029	0.019	0.020	0.020	0.019	0.020	0.021	0.019	0.019	0.019	0.021	0.019	0.019	0.020	0.019	0.019	0.020	0.018	0.020	0.020	0.019	0.015	0.019	0.019	0.019	0.020	0.019	0.019	0.019	0.019	0.018	0.020	
	2回目	0.029	0.019	0.020	0.020	0.019	0.020	0.021	0.019	0.020	0.019	0.021	0.019	0.019	0.020	0.019	0.019	0.021	0.018	0.020	0.020	0.019	0.014	0.019	0.019	0.022	0.020	0.019	0.019	0.019	0.019	0.018	0.019	
	3回目	0.030	0.019	0.019	0.019	0.019	0.020	0.021	0.019	0.020	0.019	0.021	0.019	0.019	0.020	0.018	0.017	0.019	0.019	0.020	0.020	0.019	0.019	0.014	0.019	0.020	0.020	0.019	0.019	0.020	0.018	0.018	0.020	
平均値(mg/L)①	0.0293	0.0190	0.0197	0.0197	0.0190	0.0200	0.0210	0.0190	0.0197	0.0190	0.0210	0.0190	0.0190	0.0200	0.0187	0.0183	0.0200	0.0180	0.0200	0.0197	0.0190	0.0143	0.0190	0.0193	0.0203	0.0200	0.0190	0.0190	0.0193	0.0187	0.0180	0.0197		
標準偏差②	0.000577	0.000000	0.000577	0.000577	0.000000	0.000000	0.000000	0.000000	0.000577	0.000000	0.000000	0.000000	0.000000	0.000000	0.000577	0.001155	0.001000	0.000000	0.000000	0.000577	0.000000	0.000577	0.000000	0.000577	0.001528	0.000000	0.000000	0.000577	0.000000	0.000577	0.000000	0.000577		
CV(室内)(%) (②/①)	1.97	0.00	2.93	2.93	0.00	0.00	0.00	0.00	2.93	0.00	0.00	0.00	0.00	0.00	3.09	6.31	5.00	0.00	0.00	2.93	0.00	4.04	0.00	2.99	7.53	0.00	0.00	2.99	3.09	0.00	2.93			
参加者平均との差	0.0099	-0.0004	0.0003	0.0003	-0.0004	0.0006	0.0016	-0.0004	0.0003	-0.0004	0.0016	-0.0004	-0.0004	0.0006	-0.0007	-0.0011	0.0006	-0.0014	0.0006	0.0003	-0.0004	-0.0051	-0.0004	-0.0001	0.0009	0.0006	-0.0004	-0.0004	-0.0001	-0.0007	-0.0014	0.0003		
Zスコア	13.378	-0.541	0.405	0.405	-0.541	0.811	2.162	-0.541	0.405	-0.541	2.162	-0.541	-0.541	0.811	-0.946	-1.486	0.811	-1.892	0.811	0.405	-0.541	-6.892	-0.541	-0.135	1.216	0.811	-0.541	-0.541	-0.135	-0.946	-1.892	0.405		
Grubbs検定による棄却	○																					○												

平均値	0.0194
標準偏差	0.00074
CV(室内)%	1.521

Zスコア	事業所数	32
$ Z \leq 2$		28
$2 < Z < 3$		2
$3 \leq Z $		2